

Meridian Medical Technologies Has Voluntarily Recalled Certain EpiPen® Auto-Injectors

Parents/Guardians,

The U.S. Food and Drug Administration is alerting consumers to Meridian Medical Technologies' voluntary recall of 13 lots of Mylan's EpiPen and EpiPen Jr (epinephrine injection). This recall is due to the potential of the devices' failure to activate. Recall lot numbers are listed below. If your child has an EpiPen Auto Injector, please follow the below steps:

STEP 1: Check the lot number on your carton or device to see if your EpiPen® Auto-Injector is affected by the recall. Consumers, who have EpiPens from lots that are not included in this recall, do not need to replace their EpiPen prior to its expiration date.

STEP 2: If your EpiPen® Auto-Injector has been recalled, contact Stericycle at 877-650-3494 to obtain a voucher code for your free replacement product. Stericycle also will provide you with a pre-paid return package to ship the product back to Stericycle.

STEP 3: Visit your pharmacy with your voucher information to redeem your free replacement.

STEP 4: Send your recalled product to Stericycle. **Do not return any devices affected by the recall until you have your replacement in hand.**

Recalled Lot Numbers:

Product/Dosage	NDC Number	Lot Number	Expiration Date
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	5GN767	April 2017
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	5GN773	April 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	5GM631	April 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	5GM640	May 2017
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	6GN215	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM082	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM072	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM081	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM088	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM199	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM091	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM198	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM087	October 2017

Consumers should contact Mylan at 800-796-9526 or customer.service@mylan.com with any questions. Stericycle's hours of operation are Monday-Friday 8 a.m.-10 p.m. ET, and Saturday and Sunday 8 a.m.-5 p.m. ET. We are aware that callers have experienced extended wait times and sincerely apologize for the inconvenience that this has caused.