

Epinephrine Auto-Injector Recall

There has been a voluntary recall of 13 lots of Mylan EpiPens® and EpiPen® Jrs (epinephrine auto-injectors). According to the Food and Drug Administration, this recall is due to the potential that these devices may contain a defective part that may result in the devices failure to activate, resulting in the improper release of medication. This voluntary recall is intended for auto-injectors distributed between December 17, 2015, and July 1, 2016. **The Office of the State Superintendent of Education is requiring that all public and public charter schools notify us if any of their epinephrine auto-injectors match the lot numbers below by sending an e-mail to osse.epi@dc.gov as soon as possible.**

School nurses are checking the auto-injectors prescribed to specific students. As a reminder, the undesignated epinephrine auto-injectors should be unpacked from their boxes and mounted. Attached is a communication to distribute to parents.

Recalled Lot Numbers:

| Product/Dosage | NDC Number | Lot Number | Expiration Date |
|----------------------------------|-------------------|-------------------|------------------------|
| EpiPen Jr Auto-Injector, 0.15 mg | 49502-501-02 | 5GN767 | April 2017 |
| EpiPen Jr Auto-Injector, 0.15 mg | 49502-501-02 | 5GN773 | April 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 5GM631 | April 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 5GM640 | May 2017 |
| EpiPen Jr Auto-Injector, 0.15 mg | 49502-501-02 | 6GN215 | September 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM082 | September 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM072 | September 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM081 | September 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM088 | October 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM199 | October 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM091 | October 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM198 | October 2017 |
| EpiPen Auto-Injector, 0.3 mg | 49502-500-02 | 6GM087 | October 2017 |